



**NOBAMED Paul Danz AG**

# UNTERSUCHUNGSFINGERLING

**REF 903460**

## Product Description and Purpose

The light-coloured, rolled, powder-free examination fingerstall is made of latex and serves for rectal examination. The thickness is  $\geq 0,15$  mm at the sheath and  $\geq 0,2$  mm at the fingertip

## Composition

Latex

No thiurames, toluenes or ureas are used.

## Normative and Legal Requirements

UNTERSUCHUNGSFINGERLING (examination fingerstall) is a medical product according to directive MDD 93/42/EEC and the regulation MDR (EU) 2017/745 and is classified as class I, rule 5.

The virus imperviousness is tested in accordance with **ASTM F1671-07**.

The product does not contain dangerous toxic substances according to REACH. It has CE marking and DIN EN ISO 15223-1 labels on all its packaging.

## Packaging


Primary packaging: PP bag

Secondary packaging: carton made of cellulose

## Storage

To be stored in a dry and dust-free environment, protected against direct sunlight.

The product bears following labels and symbols:



Date of information: April 01, 2020 [REV 4]  
replaced: September 25, 2015 [REV 3]